

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

## Customer PO # :

| Packaged Product: 8643         |                          |                        |          |
|--------------------------------|--------------------------|------------------------|----------|
| . aonagoa oaaon                | A HYDROXIDE 0 1667 NORMA | ZDROXIDE 0.1667 NORMAL |          |
| Test                           | Target/UOM               | Range                  | Result   |
| PREPARED TO FORMULATION ON     | YES                      | naliye                 | YES      |
| FILE                           | YES                      |                        | 120      |
| Sample ID                      | 120                      |                        |          |
| Tested 08/06/20                |                          |                        |          |
| Tester EDGA                    |                          |                        |          |
| Recorded 08/06/20              |                          |                        |          |
| SAMPLE COLLECTION & HANDLING   | i YES                    |                        | YES      |
| PER LAB040                     | YES                      |                        |          |
| Sample ID                      |                          |                        |          |
| Tested 08/06/20                |                          |                        |          |
| Tester EDGA                    |                          |                        |          |
| Recorded 08/06/20              |                          |                        |          |
| DATE OF SAMPLING FOR TEST      | MM/DD/YYYY               |                        | 08/06/20 |
| PURPOSES                       | MM/DD/YYYY               |                        |          |
| Sample ID                      |                          |                        |          |
| Tested 08/06/20                |                          |                        |          |
| Tester EDGA                    |                          |                        |          |
| Recorded 08/06/20              |                          |                        |          |
| AMBIENT TEMPERATURE DURING     | 25.0                     | 10.0 - 35.0            | 21.1     |
| SAMPLING DEGC                  | DEG C                    |                        |          |
| Sample ID                      |                          |                        |          |
| Tested 08/06/20                |                          |                        |          |
| Tester EDGA                    |                          |                        |          |
| Recorded 08/06/20              |                          |                        |          |
| DATE OF TESTING OF SAMPLE      | MM/DD/YYYY               |                        | 08/06/20 |
|                                | MM/DD/YYYY               |                        |          |
| Sample ID                      |                          |                        |          |
| Tested 08/06/20                |                          |                        |          |
| Tester EDGA                    |                          |                        |          |
| Recorded 08/06/20              |                          |                        |          |
| AMBIENT TEST TEMPERATURE       | 20.0                     | 15.0 - 30.0            | 21.1     |
| DEG C                          | DEG C                    |                        |          |
| Sample ID                      |                          |                        |          |
| Tested 08/06/20                |                          |                        |          |
| Tester EDGA                    |                          |                        |          |
| Recorded 08/06/20              | 0.1007                   | 0.1664 - 0.1670        | 0.1667   |
| NORMALITY TO 4 DEC PLACES      |                          | 0.1664 - 0.1670        | 0.1667   |
| Comula ID                      | EQUIV/LITER              |                        |          |
| Sample ID<br>Tested 08/06/20   |                          |                        |          |
| Tested 08/06/20<br>Tester EDGA |                          |                        |          |
| Recorded 08/06/20              |                          |                        |          |
| QCSP7: STANDARDIZATION OF      | YES                      |                        | YES      |
| AQUEOUS BASES                  | METHOD                   |                        | 120      |
| Sample ID                      |                          |                        |          |
| Tested 08/06/20                |                          |                        |          |
| Tester EDGA                    |                          |                        |          |
| Recorded 08/06/20              |                          |                        |          |
|                                |                          |                        |          |



## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

| Sample ID   |            |
|---|------------|
| Tested 08/06/20                                   |            |
| Tester EDGA                                       |            |
| Recorded 08/06/20                                 |            |
| APPEARANCE AND COLOR CLEAR LIQUID PASS            |            |
| APPEARANCE  |            |
| Sample ID   |            |
| Tested 08/06/20                                   |            |
| Tester EDGA                                       |            |
| Recorded 08/06/20                                 |            |
| INSTRUMENTS USED DURING INSTRUMENT M83,T26,M55,M9 | 1,BID82647 |
| PREPARATION                                       |            |
| Sample ID   |            |
| Tested 08/06/20                                   |            |
| Tester EDGA                                       |            |
| Recorded 08/06/20                                 |            |
| Actual Density 1.00571 1.00565 - 1.00577 1.00573  | 3          |
| GM/ML   |            |
| Sample ID   |            |
| Tested 08/06/20                                   |            |
| Tester EDGA                                       |            |
| Recorded 08/06/20                                 |            |
| Lot # 0080447                                     |            |
| Made 08/06/20                                     |            |

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

ichard For

**Quality Manager**