

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> BAY008F			
HYDROCHLORIC ACID 3.00% W/W CERTIFIED			
<b>Test</b>	<b>Target/UOM</b>	<b>Range</b>	<b>Result</b>
<b>PREPARED TO FORMULATION ON</b>	YES		YES
<b>FILE</b>	YES		
Sample ID			
Tested 08/06/20			
Tester OTAY			
Recorded 08/06/20			
<b>ASSAY OF PRODUCT, 2 DECIMAL PLACES</b>	3.00 W/W%	2.95 - 3.05	3.03
Sample ID			
Tested 08/06/20			
Tester OTAY			
Recorded 08/06/20			
<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE APPEARANCE		PASS
Sample ID			
Tested 08/06/20			
Tester OTAY			
Recorded 08/06/20			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M39,EF4-YP,M55,M81,BID084 46,TIDS82
Sample ID			
Tested 08/06/20			
Tester OTAY			
Recorded 08/06/20			
<b>EXPIRATION DATE</b>	MM/DD/YY MM/DD/YYYY		08/30/22
Sample ID			
Tested 08/06/20			
Tester OTAY			
Recorded 08/06/20			
<b>Lot #</b>	0080422		
<b>Made</b>	08/06/20		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

