



Aqua Solutions, Inc.  
 Ph: 281-479-2569  
 Fax: 281-479-2790  
 P.O. Box 70  
 Deer Park, TX 77536  
 Website: [www.aquasolutions.org](http://www.aquasolutions.org)

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

Test	Target/UOM	Range	Result
<b>Packaged Product:</b> ERL058F			
	POTASSIUM HYDROXIDE 0.01 NORMAL IN IPA-7235		
<b>PREPARED TO FORMULATION ON</b>	YES		YES
<b>FILE</b>	YES		
Sample ID			
Tested 08/19/21			
Tester DVJO			
Recorded 08/19/21			
<b>SAMPLE COLLECTION &amp; HANDLING</b>	YES		YES
<b>PER LAB040</b>	YES		
Sample ID			
Tested 08/19/21			
Tester DVJO			
Recorded 08/19/21			
<b>DATE OF SAMPLING FOR TEST</b>	MM/DD/YYYY		08/19/21
<b>PURPOSES</b>	MM/DD/YYYY		
Sample ID			
Tested 08/19/21			
Tester DVJO			
Recorded 08/19/21			
<b>AMBIENT TEMPERATURE DURING</b>	25.0	10.0 - 35.0	22.4
<b>SAMPLING DEGC</b>	DEG C		
Sample ID			
Tested 08/19/21			
Tester DVJO			
Recorded 08/19/21			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY		08/19/21
	MM/DD/YYYY		
Sample ID			
Tested 08/19/21			
Tester DVJO			
Recorded 08/19/21			
<b>AMBIENT TEST TEMPERATURE</b>	20.0	15.0 - 30.0	22.4
<b>DEG C</b>	DEG C		
Sample ID			
Tested 08/19/21			
Tester DVJO			
Recorded 08/19/21			
<b>LOT NUMBER OF ORIGINAL</b>	LOT #		1011113
<b>MATERIAL</b>			
Sample ID			
Tested 08/19/21			
Tester DVJO			
Recorded 08/19/21			
<b>Actual Normality</b>	0.0100	0.0094 - 0.0106	0.0094
	EQ/L		
Sample ID			
Tested 08/19/21			
Tester DVJO			
Recorded 08/19/21			



Aqua Solutions, Inc.  
 Ph: 281-479-2569  
 Fax: 281-479-2790  
 P.O. Box 70  
 Deer Park, TX 77536  
 Website: [www.aquasolutions.org](http://www.aquasolutions.org)

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Actual Density</b>	0.78630	0.78512 - 0.78748	0.78781
	GM/ML		
Sample ID			
Tested	08/19/21		
Tester	DVJO		
Recorded	08/19/21		
<b>N.I.S.T. Traceable to SRM 84</b>	YES		YES
	YES		
Sample ID			
Tested	08/19/21		
Tester	DVJO		
Recorded	08/19/21		
<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested	08/19/21		
Tester	DVJO		
Recorded	08/19/21		
<b>INSTRUMENTS USED DURING PREPARATION</b>	LAB EQUIPMENT		M55,BID02271,PIDM81
	INSTRUMENT		
Sample ID			
Tested	08/19/21		
Tester	DVJO		
Recorded	08/19/21		
<b>EXPIRATION DATE</b>	MM/DD/YYYY		08/31/21
	MM/DD/YYYY		
Sample ID			
Tested	08/19/21		
Tester	DVJO		
Recorded	08/19/21		
<b>SHELF LIFE OF PRODUCT</b>	1 WEEK		1 WEEK
	TIME		
Sample ID			
Tested	08/19/21		
Tester	DVJO		
Recorded	08/19/21		
<b>Lot #</b>	1081606		
<b>Made</b>	08/19/21		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.