



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 7350			
POTASSIUM IODIDE-IODATE 0.1 NORMAL			
<b>Test</b>	<b>Target/UOM</b>	<b>Range</b>	<b>Result</b>
<b>Actual Normality</b>	0.1000 EQ/L	0.0995 - 0.1005	0.1003
Sample ID			
Tested	09/03/21		
Tester	DHOL		
Recorded	09/03/21		
<b>Actual Density</b>	1.02697 GM/ML	1.02568 - 1.02825	1.02691
Sample ID			
Tested	09/03/21		
Tester	DHOL		
Recorded	09/03/21		
<b>QCSP33:POTASSIUM IODATE-IODIDE TITRANT</b>	YES QCSP33		YES
Sample ID			
Tested	09/03/21		
Tester	DHOL		
Recorded	09/03/21		
<b>N.I.S.T. Traceable to SRM 136</b>	Yes YES		YES
Sample ID			
Tested	09/03/21		
Tester	DTRA		
Recorded	09/03/21		
<b>PREPARED TO FORMULATION ON FILE</b>	YES YES		YES
Sample ID			
Tested	09/03/21		
Tester	DHOL		
Recorded	09/03/21		
<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	09/03/21		
Tester	DHOL		
Recorded	09/03/21		
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M91,M83,M55,P314,(O)T-225
Sample ID			
Tested	09/03/21		
Tester	DHOL		
Recorded	09/03/21		
<b>SHELF LIFE OF PRODUCT</b>	1 YEAR TIME		1 YEAR
Sample ID			
Tested	09/03/21		
Tester	DHOL		
Recorded	09/03/21		



Aqua Solutions, Inc.  
 Ph: 281-479-2569  
 Fax: 281-479-2790  
 P.O. Box 70  
 Deer Park, TX 77536  
 Website: [www.aquasolutions.org](http://www.aquasolutions.org)

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>EXPIRATION DATE</b>	MM/DD/YYYY	09/30/22
	MM/DD/YYYY	
Sample ID		
Tested	09/03/21	
Tester	DHOL	
Recorded	09/03/21	
<b>Lot #</b>	1082526	
<b>Made</b>	09/03/21	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**