



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> HOU043F			
Aromatics by HPLC Quality Control Standard ASTM D6591			
<b>Test</b>	<b>Target/UOM</b>	<b>Range</b>	<b>Result</b>
<b>PREPARED TO FORMULATION ON</b>	YES		YES
<b>FILE</b>	YES		
Sample ID			
Tested 09/01/21			
Tester ADAV			
Recorded 09/01/21			
<b>CALCULATED GRAMS</b>	50.0000	49.9900 - 50.0100	50.0000
<b>CYCLOHEXANE/100ML</b>	GM/100ML		
Sample ID			
Tested 09/01/21			
Tester ADAV			
Recorded 09/01/21			
<b>CALCULATED GRAMS</b>	15.0000	14.9900 - 15.0100	15.0004
<b>O-XYLENE/100ML</b>	GM/100ML		
Sample ID			
Tested 09/01/21			
Tester ADAV			
Recorded 09/01/21			
<b>CALCULATED GRAMS</b>	0.5000	0.4800 - 0.5200	0.5002
<b>1-METHYLNAPHTHALENE/100M</b>	GM/100ML		
Sample ID			
Tested 09/01/21			
Tester ADAV			
Recorded 09/01/21			
<b>GM/100ML PHENANTHRENE (P1943)</b>	0.3000	0.2800 - 0.3200	0.3004
	GRAMS		
Sample ID			
Tested 09/01/21			
Tester ADAV			
Recorded 09/01/21			
<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested 09/01/21			
Tester ADAV			
Recorded 09/01/21			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M21,VF100-F5
Sample ID			
Tested 09/01/21			
Tester ADAV			
Recorded 09/01/21			
<b>EXPIRATION DATE</b>	MM/DD/YY		09/30/22
	MM/DD/YYYY		
Sample ID			
Tested 09/01/21			
Tester ADAV			
Recorded 09/01/21			



Aqua Solutions, Inc.  
 Ph: 281-479-2569  
 Fax: 281-479-2790  
 P.O. Box 70  
 Deer Park, TX 77536  
 Website: [www.aquasolutions.org](http://www.aquasolutions.org)

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>SHELF LIFE OF PRODUCT</b>	1 YEAR	1 YEAR
	TIME	
Sample ID		
Tested	09/01/21	
Tester	ADAV	
Recorded	09/01/21	
<b>Lot #</b>	1082755	
<b>Made</b>	09/01/21	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**