



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: SAY001F			
1.0 mg/L 14 Component Mixed Metal Working Solution			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 10/25/21			
Tester MMAT			
Recorded 10/25/21			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 10/25/21			
Tester MMAT			
Recorded 10/25/21			
CALCULATED ALUMINUM MG/L NIST	1.00 MG/L	0.95 - 1.05	1.00
Sample ID			
Tested 10/25/21			
Tester MMAT			
Recorded 10/25/21			
CALCULATED MG/L SILICON	1.00 MG/L	0.95 - 1.05	1.00
Sample ID			
Tested 10/25/21			
Tester MMAT			
Recorded 10/25/21			
CALCULATED VANADIUM MG/L NIST	1.00 MG/L	0.95 - 1.05	1.00
Sample ID			
Tested 10/25/21			
Tester MMAT			
Recorded 10/25/21			
CALCULATED NICKEL MG/L NIST	1.00 MG/L	0.95 - 1.05	1.00
Sample ID			
Tested 10/25/21			
Tester MMAT			
Recorded 10/25/21			
CALCULATED IRON MG/L NIST	1.00 MG/L	0.95 - 1.05	1.00
Sample ID			
Tested 10/25/21			
Tester MMAT			
Recorded 10/25/21			
CALCULATED ALUMINUM MG/L NIST	1.00 MG/L	0.95 - 1.05	1.00
Sample ID			
Tested 10/25/21			
Tester MMAT			
Recorded 10/25/21			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

CALCULATED CALCIUM MG/L NIST	1.00	0.95 - 1.05	1.00
	MG/L		
Sample ID			
Tested	10/25/21		
Tester	MMAT		
Recorded	10/25/21		
CALCULATED PHOSPHORUS MG/L NIST	1.00	0.95 - 1.05	1.00
	MG/L		
Sample ID			
Tested	10/25/21		
Tester	MMAT		
Recorded	10/25/21		
CALCULATED ALUMINUM MG/L NIST	1.00	0.95 - 1.05	1.00
	MG/L		
Sample ID			
Tested	10/25/21		
Tester	MMAT		
Recorded	10/25/21		
CALCULATED POTASSIUM MG/L NIST	1.00	0.95 - 1.05	1.00
	MG/L		
Sample ID			
Tested	10/25/21		
Tester	MMAT		
Recorded	10/25/21		
CALCULATED LEAD MG/L NIST	1.00	0.95 - 1.05	1.00
	MG/L		
Sample ID			
Tested	10/25/21		
Tester	MMAT		
Recorded	10/25/21		
CALCULATED MAGNESIUM MG/L NIST	1.00	0.95 - 1.05	1.00
	MG/L		
Sample ID			
Tested	10/25/21		
Tester	MMAT		
Recorded	10/25/21		
CALCULATED MG/L ARSENIC (AS)	1.00	0.95 - 1.05	1.00
	MG/L		
Sample ID			
Tested	10/25/21		
Tester	MMAT		
Recorded	10/25/21		
CALCULATED MG/L COPPER (CU)	1.00	0.95 - 1.05	1.00
	MG/L		
Sample ID			
Tested	10/25/21		
Tester	MMAT		
Recorded	10/25/21		



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	M21,VF500-X27,P150,P341,BOTTLE
Sample ID Tested 10/25/21 Tester MMAT Recorded 10/25/21		
EXPIRATION DATE Sample ID Tested 10/25/21 Tester MMAT Recorded 10/25/21	MM/DD/YYYY MM/DD/YYYY	10/30/22
Lot # 1102551 Made 10/25/21		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager