



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 6600			
PERCHLORIC ACID TITRANT 0.1 NORMAL ACETOUS			
Test	Target/UOM	Range	Result
Actual Normality	0.1000 EQ/L	0.0995 - 0.1005	0.0998
Sample ID			
Tested 12/08/21			
Tester MDON			
Recorded 12/08/21			
QCSP16:STANDARDIZATION OF ACETOUS PERCHL	YES QCSP16		YES
Sample ID			
Tested 12/08/21			
Tester MDON			
Recorded 12/08/21			
N.I.S.T. Traceable to SRM 84	Yes YES		YES
Sample ID			
Tested 12/08/21			
Tester MDON			
Recorded 12/08/21			
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested 12/08/21			
Tester MDON			
Recorded 12/08/21			
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested 12/08/21			
Tester MDON			
Recorded 12/08/21			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M83,T-155,EF6-AD,M55,M105, M91
Sample ID			
Tested 12/08/21			
Tester MDON			
Recorded 12/08/21			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		12/30/22
Sample ID			
Tested 12/08/21			
Tester MDON			
Recorded 12/08/21			
%WATER	0.020-0.050 %WATER		267
Sample ID			
Tested 12/08/21			
Tester MDON			
Recorded 12/08/21			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

Actual Density	1.05793	1.05761 - 1.05825	1.05798
	GM/ML		
Sample ID			
Tested	12/08/21		
Tester	MDON		
Recorded	12/08/21		
AMENDMENT TO LOT#:	LOT # HAS BEEN REVISED		1111856
	LOT NUMBER		
Sample ID			
Tested	12/08/21		
Tester	MDON		
Recorded	12/08/21		
Lot #	1111856		
Made	12/08/21		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager