



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8410			
SILVER NITRATE 0.0282 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 12/07/21			
Tester MDON			
Recorded 12/07/21			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 12/07/21			
Tester MDON			
Recorded 12/07/21			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		12/07/21
Sample ID			
Tested 12/07/21			
Tester MDON			
Recorded 12/07/21			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	17.4
Sample ID			
Tested 12/07/21			
Tester MDON			
Recorded 12/07/21			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		12/07/21
Sample ID			
Tested 12/07/21			
Tester MDON			
Recorded 12/07/21			
AMBIENT TEST TEMPERATURE DEG C	20.0 DEG C	15.0 - 30.0	17.4
Sample ID			
Tested 12/07/21			
Tester MDON			
Recorded 12/07/21			
Actual Normality	0.0282 EQ/L	0.0279 - 0.0285	0.0281
Sample ID			
Tested 12/07/21			
Tester MDON			
Recorded 12/07/21			
QCSP54:STANDARDIZATION OF LOW NORMAL SN	YES QCSP54		YES
Sample ID			
Tested 12/07/21			
Tester MDON			
Recorded 12/07/21			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

N.I.S.T. Traceable to SRM 919	YES		YES
	YES		
Sample ID			
Tested	12/07/21		
Tester	MDON		
Recorded	12/07/21		
Actual Density	1.00225	1.00203 - 1.00247	1.00221
	GM/ML		
Sample ID			
Tested	12/07/21		
Tester	MDON		
Recorded	12/07/21		
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested	12/07/21		
Tester	MDON		
Recorded	12/07/21		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M21,T-42,M55,M91
Sample ID			
Tested	12/07/21		
Tester	MDON		
Recorded	12/07/21		
Lot #	1112972		
Made	12/07/21		

Quality Manager