



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: VUL380F			
SODIUM HYDROXIDE 0.001 NORMAL IN METHANOL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 04/21/22			
Tester ADAV			
Recorded 04/21/22			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 04/21/22			
Tester ADAV			
Recorded 04/21/22			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		04/20/22
Sample ID			
Tested 04/21/22			
Tester ADAV			
Recorded 04/21/22			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	21.3
Sample ID			
Tested 04/21/22			
Tester ADAV			
Recorded 04/21/22			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		04/21/22
Sample ID			
Tested 04/21/22			
Tester ADAV			
Recorded 04/21/22			
AMBIENT TEST TEMPERATURE DEGC	20.0 DEG C	15.0 - 30.0	21.3
Sample ID			
Tested 04/21/22			
Tester ADAV			
Recorded 04/21/22			
NORMALITY TO 5 DECIMAL PLACES	0.00100 EQUIV/LITER	0.00090 - 0.00120	0.00111
Sample ID			
Tested 04/21/22			
Tester ADAV			
Recorded 04/21/22			
Actual Density	0.79157 GM/ML	0.79066 - 0.79249	0.79185
Sample ID			
Tested 04/21/22			
Tester ADAV			
Recorded 04/21/22			




Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE	PASS
Sample ID		
Tested	04/21/22	
Tester	ADAV	
Recorded	04/21/22	
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	M91,EF4-YN
Sample ID		
Tested	04/21/22	
Tester	ADAV	
Recorded	04/21/22	
EXPIRATION DATE	MM/DD/YYYY	06/26/22
	MM/DD/YYYY	
Sample ID		
Tested	04/21/22	
Tester	ADAV	
Recorded	04/21/22	
Lot #	2041844	
Made	04/21/22	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



Quality Manager