



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 4112			
HYDROCHLORIC ACID 0.02 NORMAL IN METHANOL			
Test	Target/UOM	Range	Result
Actual Normality	0.0200	0.0195 - 0.0205	0.0198
	EQ/L		
Sample ID			
Tested 05/03/22			
Tester MDON			
Recorded 05/03/22			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 05/03/22			
Tester MDON			
Recorded 05/03/22			
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		05/03/22
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 05/03/22			
Tester MDON			
Recorded 05/03/22			
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	21.3
SAMPLING DEGC	DEG C		
Sample ID			
Tested 05/03/22			
Tester MDON			
Recorded 05/03/22			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		05/03/22
	MM/DD/YYYY		
Sample ID			
Tested 05/03/22			
Tester MDON			
Recorded 05/03/22			
AMBIENT TEST TEMPERATURE	20.0	10.0 - 30.0	21.3
DEG C	DEG C		
Sample ID			
Tested 05/03/22			
Tester MDON			
Recorded 05/03/22			
QCSP5: STANDARDIZATION OF	YES		YES
AQUEOUS ACIDS	METHOD		
Sample ID			
Tested 05/03/22			
Tester MDON			
Recorded 05/03/22			
N.I.S.T. Traceable to SRM 723	Yes		YES
	YES		
Sample ID			
Tested 05/03/22			
Tester MDON			
Recorded 05/03/22			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested	05/03/22		
Tester	MDON		
Recorded	05/03/22		
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested	05/03/22		
Tester	MDON		
Recorded	05/03/22		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M103,VF2-A1,M55,M91,BID064 19,TIDS3
Sample ID			
Tested	05/03/22		
Tester	MDON		
Recorded	05/03/22		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		08/30/22
Sample ID			
Tested	05/03/22		
Tester	MDON		
Recorded	05/03/22		
SHELF LIFE OF PRODUCT	3 MONTHS TIME		3 MONTHS
Sample ID			
Tested	05/03/22		
Tester	MDON		
Recorded	05/03/22		
Actual Density	0.79274 GM/ML	0.79215 - 0.79334	0.79267
Sample ID			
Tested	05/03/22		
Tester	MDON		
Recorded	05/03/22		
Lot #	2042870		
Made	05/03/22		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 05/03/22 at 8:27 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager