



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 7235			
POTASSIUM HYDROXIDE 0.01 N. IN IPA ASTM D3242, D974			
Test	Target/UOM	Range	Result
Prepared to ASTM D3242-98	Yes		YES
	PASS		
Sample ID			
Tested 05/04/22			
Tester JENC			
Recorded 05/04/22			
Actual Normality	0.0100	0.0095 - 0.0105	0.0100
	EQ/L		
Sample ID			
Tested 05/04/22			
Tester JENC			
Recorded 05/04/22			
N.I.S.T. Traceable to SRM 84	Yes		YES
	YES		
Sample ID			
Tested 05/04/22			
Tester JENC			
Recorded 05/04/22			
PREPARED TO FORMULATION ON FILE	YES		YES
	YES		
Sample ID			
Tested 05/04/22			
Tester JENC			
Recorded 05/04/22			
APPEARANCE AND COLOR	CLEAR WATER WHITE TO LIGHT HAZY APPEARANCE		PASS
Sample ID			
Tested 05/04/22			
Tester JENC			
Recorded 05/04/22			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		T169,M55,M100,M91,BID00280 ,TIDS827
Sample ID			
Tested 05/04/22			
Tester JENC			
Recorded 05/04/22			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		05/30/23
Sample ID			
Tested 05/04/22			
Tester JENC			
Recorded 05/04/22			
Actual Density	0.78610	0.78521 - 0.78699	0.78650
	GM/ML		
Sample ID			
Tested 05/04/22			
Tester JENC			
Recorded 05/04/22			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

SHELF LIFE OF PRODUCT

SHELF LIFE
TIME

1 YEARS

Sample ID

Tested 05/04/22

Tester JENC

Recorded 05/04/22

Lot # 2042241

Made 05/04/22

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager