



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 8449			
SILVER NITRATE 0.25 NORMAL NIST TRACEABLE SOLUTION			
Test	Target/UOM	Range	Result
<b>Actual Normality</b>	0.2500	0.2495 - 0.2505	0.2497
	EQ/L		
Sample ID			
Tested 05/04/22			
Tester DTRA			
Recorded 05/04/22			
<b>SAMPLE COLLECTION &amp; HANDLING</b>	YES		YES
<b>PER LAB040</b>	YES		
Sample ID			
Tested 05/04/22			
Tester DTRA			
Recorded 05/04/22			
<b>DATE OF SAMPLING FOR TEST</b>	MM/DD/YYYY		05/03/22
<b>PURPOSES</b>	MM/DD/YYYY		
Sample ID			
Tested 05/04/22			
Tester DTRA			
Recorded 05/04/22			
<b>AMBIENT TEMPERATURE DURING</b>	25.0	10.0 - 35.0	21.3
<b>SAMPLING DEGC</b>	DEG C		
Sample ID			
Tested 05/04/22			
Tester DTRA			
Recorded 05/04/22			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY		05/03/22
	MM/DD/YYYY		
Sample ID			
Tested 05/04/22			
Tester DTRA			
Recorded 05/04/22			
<b>AMBIENT TEST TEMPERATURE</b>	20.0	15.0 - 30.0	21.3
<b>DEG C</b>	DEG C		
Sample ID			
Tested 05/04/22			
Tester DTRA			
Recorded 05/04/22			
<b>PREPARED TO FORMULATION ON</b>	YES		YES
<b>FILE</b>	YES		
Sample ID			
Tested 05/04/22			
Tester DTRA			
Recorded 05/04/22			
<b>QCSP53: SILVER NITRATE STD ON</b>	YES		YES
<b>AUTO-TITRA</b>	METHOD		
Sample ID			
Tested 05/04/22			
Tester DTRA			
Recorded 05/04/22			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE LIQUID APPEARANCE	PASS
Sample ID		
Tested	05/04/22	
Tester	DTRA	
Recorded	05/04/22	
<b>INSTRUMENTS USED DURING PREPARATION</b>	<b>INSTRUMENT</b>	T219,M55,M81,BID00787
Sample ID		
Tested	05/04/22	
Tester	DTRA	
Recorded	05/04/22	
<b>EXPIRATION DATE</b>	<b>MM/DD/YY</b> <b>MM/DD/YYYY</b>	11/30/23
Sample ID		
Tested	05/04/22	
Tester	DTRA	
Recorded	05/04/22	
<b>SHELF LIFE OF PRODUCT</b>	<b>18 MONTHS</b> <b>TIME</b>	18 MONTHS
Sample ID		
Tested	05/04/22	
Tester	DTRA	
Recorded	05/04/22	
<b>Actual Density</b>	1.03373 GM/ML	1.03339 - 1.03407 1.03347
Sample ID		
Tested	05/04/22	
Tester	DTRA	
Recorded	05/04/22	
<b>Lot #</b>	2042553	
<b>Made</b>	05/04/22	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**