



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8798			
SODIUM THIOSULFATE 0.02 NORMAL			
Test	Target/UOM	Range	Result
Actual Normality	0.0200 EQ/L	0.0195 - 0.0205	0.0203
Sample ID			
Tested 04/19/22			
Tester MDON			
Recorded 04/19/22			
QCSP10:STANDARDIZATION OF SODIUM THIOSUL	YES QCSP10		YES
Sample ID			
Tested 04/19/22			
Tester MDON			
Recorded 04/19/22			
TRACEABLE TO N.I.S.T. SRM 3191	Yes YES		YES
Sample ID			
Tested 04/19/22			
Tester MDON			
Recorded 04/19/22			
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested 04/19/22			
Tester MDON			
Recorded 04/19/22			
Actual Density	1.00095 GM/ML	1.00076 - 1.00115	1.00106
Sample ID			
Tested 04/19/22			
Tester MDON			
Recorded 04/19/22			
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested 04/19/22			
Tester MDON			
Recorded 04/19/22			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		VF6-22,M81,T-27,M55,
Sample ID			
Tested 04/19/22			
Tester MDON			
Recorded 04/19/22			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		10/30/22
Sample ID			
Tested 04/19/22			
Tester MDON			
Recorded 04/19/22			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

SHELF LIFE OF PRODUCT

**SHELF LIFE
 TIME**

6 MONTHS

Sample ID

Tested 04/19/22

Tester MDON

Recorded 04/19/22

Lot # 2040621

Made 04/19/22

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager