



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 1525-4LC
 BUFFER REFERENCE pH 10.0/BLUE (4LC)

Lot # 2050993

Chemical Product: 1525
 BUFFER REFERENCE pH 10.0/BLUE

Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 05/12/22			
Tester BPHI			
Recorded 05/12/22			
PH OF PRODUCT/SOLUTION	10.000	9.990 - 10.010	10.008
Sample ID	PH		
Tested 05/12/22			
Tester BPHI			
Recorded 05/12/22			
+ PH Measurement Uncertainty	0.014	= 0.014	0.014
Sample ID	PH		
Tested 05/18/22			
Tester SNEL			
Recorded 05/18/22			
QCSP 20 PH CALIBRATION METHODS	YES		YES
Sample ID	METHOD		
Tested 05/12/22			
Tester BPHI			
Recorded 05/12/22			
Actual Temp	25.0	24.9 - 25.1	25.0
Sample ID	DEG C		
Tested 05/12/22			
Tester BPHI			
Recorded 05/12/22			
pH Traceable to N.I.S.T. SRM'S	YES		YES
Sample ID	YES		
Tested 05/12/22			
Tester BPHI			
Recorded 05/12/22			
PH NIST TRACEABLE TO SRM 185	YES		YES
Sample ID	YES		
Tested 05/12/22			
Tester BPHI			
Recorded 05/12/22			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

NIST TRACEABLE TO SRM186I AND 186II	YES YES	YES
Sample ID		
Tested 05/12/22		
Tester BPHI		
Recorded 05/12/22		
PH NIST TRACEABLE TO SRM 187	YES YES	YES
Sample ID		
Tested 05/12/22		
Tester BPHI		
Recorded 05/12/22		
TRACEABLE TO NIST SRM 191I AND SRM 191II	YES YES	YES
Sample ID		
Tested 05/12/22		
Tester BPHI		
Recorded 05/12/22		
APPEARANCE AND COLOR	CLEAR BLUE/ODORLESS APPEARANCE	PASS
Sample ID		
Tested 05/12/22		
Tester BPHI		
Recorded 05/12/22		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	M81,M56,T-IX,
Sample ID		
Tested 05/12/22		
Tester BPHI		
Recorded 05/12/22		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY	05/30/24
Sample ID		
Tested 05/12/22		
Tester BPHI		
Recorded 05/12/22		
SHELF LIFE OF PRODUCT	2 YEARS TIME	2 YEARS
Sample ID		
Tested 05/12/22		
Tester BPHI		
Recorded 05/12/22		

Made 05/12/22



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 05/26/22 at 10:52 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager