



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 4113			
HYDROCHLORIC ACID 0.0333 NORMAL (N/30)			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 05/26/22			
Tester JENC			
Recorded 05/26/22			
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES		YES
Sample ID			
Tested 05/26/22			
Tester JENC			
Recorded 05/26/22			
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY		05/26/22
Sample ID			
Tested 05/26/22			
Tester JENC			
Recorded 05/26/22			
<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0 DEG C	10.0 - 35.0	21.0
Sample ID			
Tested 05/26/22			
Tester JENC			
Recorded 05/26/22			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY MM/DD/YYYY		05/26/22
Sample ID			
Tested 05/26/22			
Tester JENC			
Recorded 05/26/22			
<b>AMBIENT TEST TEMPERATURE DEGC</b>	20.0 DEG C	15.0 - 30.0	21.0
Sample ID			
Tested 05/26/22			
Tester JENC			
Recorded 05/26/22			
<b>Actual Normality</b>	0.0333 EQ/L	0.0328 - 0.0335	0.0331
Sample ID			
Tested 05/26/22			
Tester JENC			
Recorded 05/26/22			
<b>QCSP5: STANDARDIZATION OF AQUEOUS ACIDS</b>	YES METHOD		YES
Sample ID			
Tested 05/26/22			
Tester JENC			
Recorded 05/26/22			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID APPEARANCE	PASS
Sample ID		
Tested	05/26/22	
Tester	JENC	
Recorded	05/26/22	
<b>INSTRUMENTS USED DURING PREPARATION</b>	<b>INSTRUMENT</b>	GC,T157,M55,M91,BID02280, TIDS82796
Sample ID		
Tested	05/26/22	
Tester	JENC	
Recorded	05/26/22	
<b>N.I.S.T. Traceable to SRM 723</b>	YES	YES
	YES	
Sample ID		
Tested	05/26/22	
Tester	JENC	
Recorded	05/26/22	
<b>Lot #</b>	2052065	
<b>Made</b>	05/26/22	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



**Quality Manager**