



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 1501			
BUFFER REFERENCE pH 7.0/CLEAR			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 05/25/22			
Tester BPHI			
Recorded 05/25/22			
QCSP 20 PH CALIBRATION METHODS	YES		YES
Sample ID	METHOD		
Tested 05/25/22			
Tester BPHI			
Recorded 05/25/22			
Actual Temp	25.0	24.0 - 26.0	25.0
	DEG C		
Sample ID			
Tested 05/25/22			
Tester BPHI			
Recorded 05/25/22			
pH Traceable to N.I.S.T. SRM'S	YES		YES
	YES		
Sample ID			
Tested 05/25/22			
Tester BPHI			
Recorded 05/25/22			
PH NIST TRACEABLE TO SRM 185	YES		YES
	YES		
Sample ID			
Tested 05/25/22			
Tester BPHI			
Recorded 05/25/22			
NIST TRACEABLE TO SRM186I AND 186II	YES		YES
	YES		
Sample ID			
Tested 05/25/22			
Tester BPHI			
Recorded 05/25/22			
PH NIST TRACEABLE TO SRM 187	YES		YES
	YES		
Sample ID			
Tested 05/25/22			
Tester BPHI			
Recorded 05/25/22			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested 05/25/22			
Tester BPHI			
Recorded 05/25/22			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

INSTRUMENTS USED DURING PREPARATION	LAB EQUIPMENT INSTRUMENT	M85,T-IX,M56
Sample ID		
Tested 05/25/22		
Tester BPHI		
Recorded 05/25/22		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY	05/30/24
Sample ID		
Tested 05/25/22		
Tester BPHI		
Recorded 05/25/22		
SHELF LIFE OF PRODUCT	SHELF LIFE TIME	2 YEARS
Sample ID		
Tested 05/25/22		
Tester BPHI		
Recorded 05/25/22		
PH 3 DECIMAL PLACES	7.000 PH	6.990 - 7.010 7.009
Sample ID		
Tested 05/25/22		
Tester BPHI		
Recorded 05/25/22		
Lot # 2051918		
Made 05/25/22		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.


Quality Manager