



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8798			
SODIUM THIOSULFATE 0.02 NORMAL			
Test	Target/UOM	Range	Result
Actual Normality	0.0200 EQ/L	0.0195 - 0.0205	0.0202
Sample ID Tested 06/08/22 Tester CNOL Recorded 06/08/22			
QCSP10:STANDARDIZATION OF SODIUM THIOSUL	YES QCSP10		YES
Sample ID Tested 06/08/22 Tester CNOL Recorded 06/08/22			
TRACEABLE TO N.I.S.T. SRM 3191	Yes YES		YES
Sample ID Tested 06/08/22 Tester CNOL Recorded 06/08/22			
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID Tested 06/08/22 Tester CNOL Recorded 06/08/22			
Actual Density	1.00095 GM/ML	1.00076 - 1.00115	1.00105
Sample ID Tested 06/08/22 Tester CNOL Recorded 06/08/22			
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID Tested 06/08/22 Tester CNOL Recorded 06/08/22			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M55,M103,T27,P359,BID06969
Sample ID Tested 06/08/22 Tester CNOL Recorded 06/08/22			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		12/30/22
Sample ID Tested 06/08/22 Tester CNOL Recorded 06/08/22			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

SHELF LIFE OF PRODUCT	SHELF LIFE TIME	6 MONTHS
Sample ID		
Tested	06/08/22	
Tester	CNOL	
Recorded	06/08/22	
Lot #	2052456	
Made	06/08/22	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager