



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: LY396F			
ORGANIC ACID CALIBRATION STOCK 1,000 PPM EA			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 06/15/22			
Tester ELAT			
Recorded 06/15/22			
CALCULATED PPM ACETATE	1000.00 PPM	995.00 - 1005.00	1000.03
Sample ID			
Tested 06/15/22			
Tester ELAT			
Recorded 06/15/22			
CALCULATED PPM FORMATE	1000.00 PPM	995.00 - 1005.00	1000.02
Sample ID			
Tested 06/15/22			
Tester ELAT			
Recorded 06/15/22			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 06/15/22			
Tester ELAT			
Recorded 06/15/22			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M99,VF1-Z21
Sample ID			
Tested 06/15/22			
Tester ELAT			
Recorded 06/15/22			
CALCULATED PPM PROPIONATE	1000.00 PPM	995.00 - 1005.00	1000.07
Sample ID			
Tested 06/16/22			
Tester ELAT			
Recorded 06/16/22			
CALCULATED PPM BUTYRATE	1000.00 PPM	995.00 - 1005.00	1000.02
Sample ID			
Tested 06/16/22			
Tester ELAT			
Recorded 06/16/22			
CALCULATED PPM CARBONATE	1000.00 PPM	995.00 - 1005.00	1000.00
Sample ID			
Tested 06/16/22			
Tester ELAT			
Recorded 06/16/22			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 06/16/22 at 6:55 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

CALCULATED PPM NITRATE	1000.00	995.00 - 1005.00	1000.05
	PPM		
Sample ID			
Tested	06/16/22		
Tester	ELAT		
Recorded	06/16/22		
Lot #	2061469		
Made	06/15/22		

Quality Manager