



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> FIS107F			
Conductivity Standard 74.0 umhos @25°C			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON</b>	YES		YES
<b>FILE</b>	YES		
Sample ID			
Tested 06/24/22			
Tester RIRO			
Recorded 06/24/22			
<b>CONDUCTIVITY @25 C.</b>	74	73 - 75	74
	US/CM		
Sample ID			
Tested 06/24/22			
Tester RIRO			
Recorded 06/24/22			
<b>QCSP30: CONDUCTIVITY</b>	YES		YES
<b>STD&amp;MEASUREMENT</b>	QCSP30		
Sample ID			
Tested 06/24/22			
Tester RIRO			
Recorded 06/24/22			
<b>TRACEABLE TO N.I.S.T. SRM 3191</b>	YES		YES
	YES		
Sample ID			
Tested 06/24/22			
Tester RIRO			
Recorded 06/24/22			
<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE		PASS
	APPEARANCE		
Sample ID			
Tested 06/24/22			
Tester RIRO			
Recorded 06/24/22			
<b>INSTRUMENTS USED DURING</b>	INSTRUMENT		T71,M92
<b>PREPARATION</b>			
Sample ID			
Tested 06/24/22			
Tester RIRO			
Recorded 06/24/22			
<b>EXPIRATION DATE</b>	MM/DD/YYYY		06/30/23
	MM/DD/YYYY		
Sample ID			
Tested 06/24/22			
Tester RIRO			
Recorded 06/24/22			
<b>Lot #</b>	2060910		
<b>Made</b>	06/24/22		



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 2

Date: 06/24/22 at 6:37 AM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

**Quality Manager**