



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 4360			
HYDROCHLORIC ACID 5.0 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 06/23/22			
Tester DTRA			
Recorded 06/23/22			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 06/23/22			
Tester DTRA			
Recorded 06/23/22			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		06/22/22
Sample ID			
Tested 06/23/22			
Tester DTRA			
Recorded 06/23/22			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	21.6
Sample ID			
Tested 06/23/22			
Tester DTRA			
Recorded 06/23/22			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		06/22/22
Sample ID			
Tested 06/23/22			
Tester DTRA			
Recorded 06/23/22			
AMBIENT TEST TEMPERATURE DEGC	20.0 DEG C	15.0 - 30.0	21.6
Sample ID			
Tested 06/23/22			
Tester DTRA			
Recorded 06/23/22			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 06/23/22			
Tester DTRA			
Recorded 06/23/22			
QCSP5: STANDARDIZATION OF AQUEOUS ACIDS	YES METHOD		YES
Sample ID			
Tested 06/23/22			
Tester DTRA			
Recorded 06/23/22			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	T49,M55,M81,BID02871,TIDS3
Sample ID		1424
Tested 06/23/22		
Tester DTRA		
Recorded 06/23/22		
NORMALITY TO 4 DEC PLACES	5.0000	4.9800 - 5.0200
	EQUIV/LITER	5.0010
Sample ID		
Tested 06/23/22		
Tester DTRA		
Recorded 06/23/22		
EXPIRATION DATE	MM/DD/YY	06/30/24
	MM/DD/YYYY	
Sample ID		
Tested 06/23/22		
Tester DTRA		
Recorded 06/23/22		
COLOR	< 10	<10
	PT-CO	
Sample ID		
Tested 06/23/22		
Tester DTRA		
Recorded 06/23/22		
N.I.S.T. Traceable to SRM 723	YES	YES
	YES	
Sample ID		
Tested 06/23/22		
Tester DTRA		
Recorded 06/23/22		
SHELF LIFE OF PRODUCT	2 YEARS	2 YEARS
	TIME	
Sample ID		
Tested 06/23/22		
Tester DTRA		
Recorded 06/23/22		
Actual Density	1.08138	1.08071 - 1.08204
	GM/ML	1.08157
Sample ID		
Tested 06/23/22		
Tester DTRA		
Recorded 06/23/22		

Lot # 2060944

Made 06/23/22



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 06/23/22 at 6:15 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager