



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: FIS108F			
Conductivity Standard 149.6 umhos @25°C			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 06/24/22			
Tester RIRO			
Recorded 06/24/22			
CONDUCTIVITY 1 DEC	149.6	149.5 - 149.7	149.6
Sample ID			
Tested 06/24/22			
Tester RIRO			
Recorded 06/24/22			
QCSP30: CONDUCTIVITY	YES		YES
STD&MEASUREMENT	QCSP30		
Sample ID			
Tested 06/24/22			
Tester RIRO			
Recorded 06/24/22			
TRACEABLE TO N.I.S.T. SRM 3191	YES		YES
	YES		
Sample ID			
Tested 06/24/22			
Tester RIRO			
Recorded 06/24/22			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	APPEARANCE		
Sample ID			
Tested 06/24/22			
Tester RIRO			
Recorded 06/24/22			
INSTRUMENTS USED DURING	INSTRUMENT		T71,M92
PREPARATION			
Sample ID			
Tested 06/24/22			
Tester RIRO			
Recorded 06/24/22			
EXPIRATION DATE	MM/DD/YYYY		06/30/23
	MM/DD/YYYY		
Sample ID			
Tested 06/24/22			
Tester RIRO			
Recorded 06/24/22			
Lot #	2060911		
Made	06/24/22		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 06/24/22 at 6:49 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Quality Manager