



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: ODP149F			
LACHAT CALIBRATION STANDARD LEVEL 6			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 06/28/22			
Tester MMAT			
Recorded 06/28/22			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 06/28/22			
Tester MMAT			
Recorded 06/28/22			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P349,M100,VF2-B17,M55,M21,
Sample ID			
Tested 06/28/22			
Tester MMAT			
Recorded 06/28/22			
CALCULATED PPM KIO3-NACL	5.0 PPM	4.9 - 5.1	5.0
Sample ID			
Tested 06/28/22			
Tester MMAT			
Recorded 06/28/22			
CALCULATED GPL SODIUM HYDROXIDE	85.00 GPL	84.96 - 85.04	85.00
Sample ID			
Tested 06/28/22			
Tester MMAT			
Recorded 06/28/22			
CALCULATED GPL SODIUM CHLORIDE	225.00 GPL	224.98 - 225.02	225.00
Sample ID			
Tested 06/28/22			
Tester MMAT			
Recorded 06/28/22			
CALCULATED GPL SODIUM CHLORATE	1.3500 GPL	1.3495 - 1.3505	1.3499
Sample ID			
Tested 06/28/22			
Tester MMAT			
Recorded 06/28/22			
ASSAY SODIUM CHLORATE	100.0 %	98.0 - 102.0	99.6
Sample ID			
Tested 06/29/22			
Tester MMAT			
Recorded 06/29/22			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

FINAL VOLUME IN LITERS	1.0	1.0 - 6.0	2.0
	LITER		
Sample ID			
Tested	06/29/22		
Tester	MMAT		
Recorded	06/29/22		
MLS OF 1000 PPM KIO3-NAOCL STOCK STD	5.0	5.0 - 30.0	10.0
	MLS		
Sample ID			
Tested	06/30/22		
Tester	MMAT		
Recorded	06/30/22		
GRAMS SODIUM HYDROXIDE	85.00	80.00 - 515.00	170.00
	GM		
Sample ID			
Tested	06/30/22		
Tester	MMAT		
Recorded	06/30/22		
GMS SODIUM CHLORIDE	225.00	220.00 - 1355.00	450.00
	GM		
Sample ID			
Tested	06/30/22		
Tester	MMAT		
Recorded	06/30/22		
GRAMS SODIUM CHLORATE	1.3500	1.3495 - 8.1005	2.7107
	GM		
Sample ID			
Tested	06/30/22		
Tester	MMAT		
Recorded	06/30/22		
Lot #	2062440		
Made	06/28/22		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager