



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
 Customer PO # :

Packaged Product: 1430			
BUFFER REFERENCE pH 1.01			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 05/06/22			
Tester DTRA			
Recorded 05/06/22			
Actual pH of product	1.010	0.990 - 1.030	1.006
	PH		
Sample ID			
Tested 05/06/22			
Tester DTRA			
Recorded 05/06/22			
QCSP 20 PH CALIBRATION METHODS	YES		YES
	METHOD		
Sample ID			
Tested 05/06/22			
Tester DTRA			
Recorded 05/06/22			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 05/06/22			
Tester DTRA			
Recorded 05/06/22			
INSTRUMENTS USED DURING PREPARATION	EQUIPMENT INSTRUMENT		M81,T-21,M98,
Sample ID			
Tested 05/06/22			
Tester DTRA			
Recorded 05/06/22			
PH NIST TRACEABLE TO SRM 189	YES		YES
	YES		
Sample ID			
Tested 05/06/22			
Tester DTRA			
Recorded 05/06/22			
PH NIST TRACEABLE TO SRM 185	YES		YES
	YES		
Sample ID			
Tested 05/06/22			
Tester DTRA			
Recorded 05/06/22			
EXPIRATION DATE	MM/DD/YY		05/30/23
	MM/DD/YYYY		
Sample ID			
Tested 05/06/22			
Tester DTRA			
Recorded 05/06/22			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

SHELF LIFE OF PRODUCT

1 YEAR
 TIME

1 YEAR

Sample ID

Tested 05/06/22

Tester DTRA

Recorded 05/06/22

Lot # 2040722

Made 05/06/22

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager