



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: LY386F			
SULFUR STANDARD 10.0 PPM W/W IN ISOOCTANE			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 06/28/22			
Tester LBYE			
Recorded 06/28/22			
Calculated ppm Sulfur	10.00	9.80 - 10.20	10.00
	PPM		
Sample ID			
Tested 06/28/22			
Tester LBYE			
Recorded 06/28/22			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 06/28/22			
Tester LBYE			
Recorded 06/28/22			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P370,VF1-Z21
Sample ID			
Tested 06/28/22			
Tester LBYE			
Recorded 06/28/22			
Lot #	2062473		
Made	06/28/22		

Quality Manager