



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: SPE509F			
TOTAL SUSPENDED SOLIDS STANDARD 1000 PPM			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 07/13/22			
Tester RIRO			
Recorded 07/13/22			
APPEARANCE AND COLOR	TURBID LIQUID APPEARANCE		PASS
Sample ID			
Tested 07/13/22			
Tester RIRO			
Recorded 07/13/22			
CALCULATED MG/L (PPM) TSS	1000.0 MG/L	995.0 - 1005.0	999.6
Sample ID			
Tested 07/13/22			
Tester RIRO			
Recorded 07/13/22			
ACT GMS CELITE	1.0000 GM	0.9990 - 1.0010	0.9996
Sample ID			
Tested 07/13/22			
Tester RIRO			
Recorded 07/13/22			
QCSP11:STANDARDIZATION OF BROMIDE-BROMAT	YES QCSP11		YES
Sample ID			
Tested 07/13/22			
Tester RIRO			
Recorded 07/13/22			
MLS OF WATER	1000 ML	995 - 1005	1000
Sample ID			
Tested 07/13/22			
Tester RIRO			
Recorded 07/13/22			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M91,VF1-26
Sample ID			
Tested 07/13/22			
Tester RIRO			
Recorded 07/13/22			
WEIGHT NIST TRACEABLE	PASS YES		PASS
Sample ID			
Tested 07/13/22			
Tester RIRO			
Recorded 07/13/22			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

EXPIRATION DATE	MM/DD/YY	07/30/23
	MM/DD/YYYY	
Sample ID		
Tested	07/13/22	
Tester	RIRO	
Recorded	07/13/22	
SHELF LIFE OF PRODUCT	SHELF LIFE	1 YEAR
	TIME	
Sample ID		
Tested	07/13/22	
Tester	RIRO	
Recorded	07/13/22	
Lot #	2070121	
Made	07/13/22	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager