



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8810			
SODIUM THIOSULFATE 0.1 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 05/10/22			
Tester MDON			
Recorded 05/10/22			
PH OF PRODUCT/SOLUTION	7.000 PH	6.000 - 12.500	10.162
Sample ID			
Tested 05/10/22			
Tester MDON			
Recorded 05/10/22			
Actual Normality	0.1000 EQ/L	0.0995 - 0.1005	0.1003
Sample ID			
Tested 05/10/22			
Tester MDON			
Recorded 05/10/22			
QCSP10:STANDARDIZATION OF SODIUM THIOSUL	YES QCSP10		YES
Sample ID			
Tested 05/10/22			
Tester MDON			
Recorded 05/10/22			
QCSP35:STANDARDIZATION OF SODIUM THIOSUL	YES QCSP35		YES
Sample ID			
Tested 05/10/22			
Tester MDON			
Recorded 05/10/22			
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested 05/10/22			
Tester MDON			
Recorded 05/10/22			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M87,M100,T-156,M98,M55,M91
Sample ID			
Tested 05/10/22			
Tester MDON			
Recorded 05/10/22			
DATE OF PREPARATION	MM/DD/YYYY		05/10/22
Sample ID			
Tested 05/10/22			
Tester MDON			
Recorded 05/10/22			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

EXPIRATION DATE	MM/DD/YY		11/30/22
	MM/DD/YYYY		
Sample ID			
Tested	05/10/22		
Tester	MDON		
Recorded	05/10/22		
SHELF LIFE	6 MONTHS		6 MONTHS
	TIME		
Sample ID			
Tested	05/10/22		
Tester	MDON		
Recorded	05/10/22		
TRACEABLE TO NIST SRM 136	YES		YES
	YES		
Sample ID			
Tested	05/10/22		
Tester	MDON		
Recorded	05/10/22		
Actual Density	1.01127	1.01080 - 1.01173	1.01143
	GM/ML		
Sample ID			
Tested	05/10/22		
Tester	MDON		
Recorded	05/10/22		
Lot #	2042908		
Made	05/10/22		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager