



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8432			
SILVER NITRATE 0.141 NORMAL			
Test	Target/UOM	Range	Result
Actual Normality	0.1410 EQ/L	0.1405 - 0.1415	0.1410
Sample ID			
Tested	07/13/22		
Tester	MDON		
Recorded	07/13/22		
+ Measurement Uncertainty	0.0022 EQ/L	= 0.0022	0.0022
Sample ID			
Tested	07/13/22		
Tester	MDON		
Recorded	07/13/22		
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested	07/13/22		
Tester	MDON		
Recorded	07/13/22		
QCSP53: SILVER NITRATE STD ON	YES		YES
AUTO-TITRA	METHOD		
Sample ID			
Tested	07/13/22		
Tester	MDON		
Recorded	07/13/22		
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		07/13/22
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested	07/13/22		
Tester	MDON		
Recorded	07/13/22		
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	26.4
SAMPLING DEGC	DEG C		
Sample ID			
Tested	07/13/22		
Tester	MDON		
Recorded	07/13/22		
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		07/13/22
	MM/DD/YYYY		
Sample ID			
Tested	07/13/22		
Tester	MDON		
Recorded	07/13/22		
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	26.4
DEG C	DEG C		
Sample ID			
Tested	07/13/22		
Tester	MDON		
Recorded	07/13/22		



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested	07/13/22		
Tester	MDON		
Recorded	07/13/22		
N.I.S.T. Traceable to SRM 919 NaCl	YES YES		YES
Sample ID			
Tested	07/13/22		
Tester	MDON		
Recorded	07/13/22		
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	07/13/22		
Tester	MDON		
Recorded	07/13/22		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M103,T230,M55,M91,B4,TIDS3 1570
Sample ID			
Tested	07/13/22		
Tester	MDON		
Recorded	07/13/22		
Actual Density	1.01826 GM/ML	1.01792 - 1.01861	1.01813
Sample ID			
Tested	07/13/22		
Tester	MDON		
Recorded	07/13/22		
Lot #	2063032		
Made	07/13/22		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager