



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: FIS034F			
Total Suspended Solids Standard 500 ppm			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 07/25/22			
Tester LBYE			
Recorded 07/25/22			
APPEARANCE AND COLOR	TURBID LIQUID APPEARANCE		PASS
Sample ID			
Tested 07/25/22			
Tester LBYE			
Recorded 07/25/22			
CALCULATED PPM	500 PPM	495 - 505	500
Sample ID			
Tested 07/25/22			
Tester LBYE			
Recorded 07/25/22			
ACT GMS CELITE	0.5000 GM	0.4995 - 0.5005	0.5004
Sample ID			
Tested 07/25/22			
Tester LBYE			
Recorded 07/25/22			
MLS OF WATER	1000 ML	995 - 1005	1000
Sample ID			
Tested 07/25/22			
Tester LBYE			
Recorded 07/25/22			
WEIGHT NIST TRACEABLE	PASS YES		PASS
Sample ID			
Tested 07/25/22			
Tester LBYE			
Recorded 07/25/22			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M21,VF1-Z10
Sample ID			
Tested 07/25/22			
Tester LBYE			
Recorded 07/25/22			
Lot #	2072196		
Made	07/25/22		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 07/25/22 at 8:02 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager