



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: CYM004F			
ACRYLAMIDE 15 PPM/ACRYLIC ACID 75 PPM			
Test	Target/UOM	Range	Result
PPM ACRYLAMIDE	15.00 PPM	14.00 - 16.00	14.98
Sample ID			
Tested	07/27/22		
Tester	LBYE		
Recorded	07/27/22		
Actual ppm Acrylic Acid	75.00 PPM	70.00 - 80.00	78.90
Sample ID			
Tested	07/27/22		
Tester	LBYE		
Recorded	07/27/22		
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested	07/27/22		
Tester	LBYE		
Recorded	07/27/22		
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested	07/27/22		
Tester	LBYE		
Recorded	07/27/22		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M21,
Sample ID			
Tested	07/27/22		
Tester	LBYE		
Recorded	07/27/22		
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY		09/27/22
Sample ID			
Tested	07/27/22		
Tester	LBYE		
Recorded	07/27/22		
SHELF LIFE OF PRODUCT	2 MONTHS TIME		2 MONTHS
Sample ID			
Tested	07/27/22		
Tester	LBYE		
Recorded	07/27/22		
Lot #	2072204		
Made	07/27/22		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 07/27/22 at 11:39 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager