



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 9021			
SULFURIC ACID 0.02 NORMAL (BCG/MR/CO3)			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID	YES		
Tested 08/08/22			
Tester MDON			
Recorded 08/08/22			
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES		YES
Sample ID	YES		
Tested 08/08/22			
Tester MDON			
Recorded 08/08/22			
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY		08/08/22
Sample ID	MM/DD/YYYY		
Tested 08/08/22			
Tester MDON			
Recorded 08/08/22			
<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0	10.0 - 35.0	27.3
Sample ID	DEG C		
Tested 08/08/22			
Tester MDON			
Recorded 08/08/22			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY		08/08/22
Sample ID	MM/DD/YYYY		
Tested 08/08/22			
Tester MDON			
Recorded 08/08/22			
<b>AMBIENT TEST TEMPERATURE DEG C</b>	20.0	15.0 - 30.0	27.3
Sample ID	DEG C		
Tested 08/08/22			
Tester MDON			
Recorded 08/08/22			
<b>NORMALITY TO 4 DEC PLACES</b>	0.0200	0.0195 - 0.0205	0.0204
Sample ID	EQUIV/LITER		
Tested 08/08/22			
Tester MDON			
Recorded 08/08/22			
<b>QCSP61A:STANDARDIZATION ACIDS USING SODI</b>	YES		YES
Sample ID	QCSP61A		
Tested 08/08/22			
Tester MDON			
Recorded 08/08/22			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE		PASS
	LIQUID APPEARANCE		
Sample ID			
Tested	08/08/22		
Tester	MDON		
Recorded	08/08/22		
<b>INSTRUMENTS USED DURING PREPARATION</b>	<b>INSTRUMENT</b>		GC,T-28,M55,BURET:06419,M91
Sample ID			
Tested	08/08/22		
Tester	MDON		
Recorded	08/08/22		
<b>Actual Density</b>	0.99899	0.99882 - 0.99916	0.99897
	GM/ML		
Sample ID			
Tested	08/08/22		
Tester	MDON		
Recorded	08/08/22		
<b>Lot #</b>	2071947		
<b>Made</b>	08/08/22		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

  
**Quality Manager**