



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

| Packaged Product: 9098 | | | |
|--|------------|-----------------|----------|
| SULFURIC ACID 5.25 NORMAL NIST TRACEABLE SOLUTION | | | |
| Test | Target/UOM | Range | Result |
| N.I.S.T. Traceable to SRM 723 | Yes | | YES |
| | YES | | |
| Sample ID | | | |
| Tested 08/11/22 | | | |
| Tester JENC | | | |
| Recorded 08/11/22 | | | |
| SAMPLE COLLECTION & HANDLING PER LAB040 | YES | | YES |
| | YES | | |
| Sample ID | | | |
| Tested 08/11/22 | | | |
| Tester JENC | | | |
| Recorded 08/11/22 | | | |
| DATE OF SAMPLING FOR TEST PURPOSES | MM/DD/YYYY | | 08/10/22 |
| | MM/DD/YYYY | | |
| Sample ID | | | |
| Tested 08/11/22 | | | |
| Tester JENC | | | |
| Recorded 08/11/22 | | | |
| AMBIENT TEMPERATURE DURING SAMPLING DEGC | 25.0 | 10.0 - 35.0 | 21.9 |
| | DEG C | | |
| Sample ID | | | |
| Tested 08/11/22 | | | |
| Tester JENC | | | |
| Recorded 08/11/22 | | | |
| DATE OF TESTING OF SAMPLE | MM/DD/YYYY | | 08/10/22 |
| | MM/DD/YYYY | | |
| Sample ID | | | |
| Tested 08/11/22 | | | |
| Tester JENC | | | |
| Recorded 08/11/22 | | | |
| AMBIENT TEST TEMPERATURE DEGC | 20.0 | 15.0 - 30.0 | 21.9 |
| | DEG C | | |
| Sample ID | | | |
| Tested 08/11/22 | | | |
| Tester JENC | | | |
| Recorded 08/11/22 | | | |
| QCSP5: STANDARDIZATION OF AQUEOUS ACIDS | YES | | 08/10/22 |
| | METHOD | | |
| Sample ID | | | |
| Tested 08/11/22 | | | |
| Tester JENC | | | |
| Recorded 08/11/22 | | | |
| Actual Normality | 5.2500 | 5.2400 - 5.2600 | 5.2476 |
| | EQ/L | | |
| Sample ID | | | |
| Tested 08/11/22 | | | |
| Tester JENC | | | |
| Recorded 08/11/22 | | | |




Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

| | | | |
|--|--|--------------------------|-------------------|
| PREPARED TO FORMULATION ON FILE | YES YES | | YES |
| Sample ID | | | |
| Tested | 08/11/22 | | |
| Tester | JENC | | |
| Recorded | 08/11/22 | | |
| APPEARANCE AND COLOR | CLEAR WATER WHITE LIQUID APPEARANCE | | PASS |
| Sample ID | | | |
| Tested | 08/11/22 | | |
| Tester | JENC | | |
| Recorded | 08/11/22 | | |
| INSTRUMENTS USED DURING PREPARATION | INSTRUMENT | | GC, VF2-A7 |
| Sample ID | | | |
| Tested | 08/11/22 | | |
| Tester | JENC | | |
| Recorded | 08/11/22 | | |
| EXPIRATION DATE | MM/DD/YYYY | | 08/30/24 |
| Sample ID | | | |
| Tested | 08/11/22 | | |
| Tester | JENC | | |
| Recorded | 08/11/22 | | |
| Actual Density | 1.15678 | 1.15583 - 1.15773 | 1.15608 |
| | GM/ML | | |
| Sample ID | | | |
| Tested | 08/11/22 | | |
| Tester | JENC | | |
| Recorded | 08/11/22 | | |
| Lot # | 2072234 | | |
| Made | 08/11/22 | | |

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



Quality Manager