



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Test	Target/UOM	Range	Result
Packaged Product: UC115F			
	SODIUM HYDROXIDE 0.5 NORMAL STD ALCOHOLIC-8647		
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 08/08/22			
Tester JENC			
Recorded 08/08/22			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 08/08/22			
Tester JENC			
Recorded 08/08/22			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		08/08/22
Sample ID			
Tested 08/08/22			
Tester JENC			
Recorded 08/08/22			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	21.6
Sample ID			
Tested 08/08/22			
Tester JENC			
Recorded 08/08/22			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		08/08/22
Sample ID			
Tested 08/08/22			
Tester JENC			
Recorded 08/08/22			
AMBIENT TEST TEMPERATURE DEGC	20.0 DEG C	15.0 - 30.0	21.6
Sample ID			
Tested 08/08/22			
Tester JENC			
Recorded 08/08/22			
Actual Normality	0.5000 EQ/L	0.4995 - 0.5005	0.5003
Sample ID			
Tested 08/08/22			
Tester JENC			
Recorded 08/08/22			
QCSP7: STANDARDIZATION OF AQUEOUS BASES	YES METHOD		YES
Sample ID			
Tested 08/08/22			
Tester JENC			
Recorded 08/08/22			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

N.I.S.T. Traceable to SRM 84 (KHP)	YES YES	YES
Sample ID		
Tested	08/08/22	
Tester	JENC	
Recorded	08/08/22	
Actual Density	0.89612 GM/ML	0.89396 - 0.89829 0.89619
Sample ID		
Tested	08/08/22	
Tester	JENC	
Recorded	08/08/22	
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE	PASS
Sample ID		
Tested	08/08/22	
Tester	JENC	
Recorded	08/08/22	
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	M83,GC,VF5-E1,M55,M91,BID 02253,TID
Sample ID		
Tested	08/08/22	
Tester	JENC	
Recorded	08/08/22	
SHELF LIFE OF PRODUCT	6 MONTHS TIME	6 MONTHS
Sample ID		
Tested	08/08/22	
Tester	JENC	
Recorded	08/08/22	
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY	02/28/23
Sample ID		
Tested	08/08/22	
Tester	JENC	
Recorded	08/08/22	
Lot #	2080344	
Made	08/08/22	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 08/08/22 at 3:11 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager