



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> EQS266F			
MIXED ACID STANDARD			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 08/18/22			
Tester MMAT			
Recorded 08/18/22			
<b>ACTUAL PPM SODIUM FORMATE</b>	1000.00 PPM	998.00 - 1002.00	999.89
Sample ID			
Tested 08/18/22			
Tester MMAT			
Recorded 08/18/22			
<b>ACTUAL PPM SODIUM ACETATE</b>	1000.00 PPM	998.00 - 1002.00	1000.03
Sample ID			
Tested 08/18/22			
Tester MMAT			
Recorded 08/18/22			
<b>ACTUAL PPM SODIUM PROPINOATE</b>	1000.00 PPM	998.00 - 1002.00	999.99
Sample ID			
Tested 08/18/22			
Tester MMAT			
Recorded 08/18/22			
<b>ACTUAL PPM SODIUM BUTYRATE</b>	1000.00 PPM	998.00 - 1002.00	1000.15
Sample ID			
Tested 08/18/22			
Tester MMAT			
Recorded 08/18/22			
<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 08/18/22			
Tester MMAT			
Recorded 08/18/22			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M21,VF1-Z2
Sample ID			
Tested 08/18/22			
Tester MMAT			
Recorded 08/18/22			
<b>SHELF LIFE OF PRODUCT</b>	1 YEAR TIME		1 YEAR
Sample ID			
Tested 08/18/22			
Tester MMAT			
Recorded 08/18/22			



Aqua Solutions, Inc.  
 Ph: 281-479-2569  
 Fax: 281-479-2790  
 P.O. Box 70  
 Deer Park, TX 77536  
 Website: [www.aquasolutions.org](http://www.aquasolutions.org)

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>EXPIRATION DATE</b>	MM/DD/YYYY	08/30/23
	MM/DD/YYYY	
Sample ID		
Tested	08/18/22	
Tester	MMAT	
Recorded	08/18/22	
<b>Lot #</b>	2080867	
<b>Made</b>	08/18/22	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**