



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Test	Target/UOM	Range	Result
Packaged Product: MOB029F			
	HYDROCHLORIC ACID 1.0 NORMAL		
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 08/24/22			
Tester JENC			
Recorded 08/24/22			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 08/24/22			
Tester JENC			
Recorded 08/24/22			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		08/23/22
Sample ID			
Tested 08/24/22			
Tester JENC			
Recorded 08/24/22			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	21.2
Sample ID			
Tested 08/24/22			
Tester JENC			
Recorded 08/24/22			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		08/23/22
Sample ID			
Tested 08/24/22			
Tester JENC			
Recorded 08/24/22			
AMBIENT TEST TEMPERATURE DEGC	20.0 DEG C	15.0 - 30.0	21.2
Sample ID			
Tested 08/24/22			
Tester JENC			
Recorded 08/24/22			
Actual Normality	1.0000 EQ/L	0.9995 - 1.0005	0.9997
Sample ID			
Tested 08/24/22			
Tester JENC			
Recorded 08/24/22			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 08/24/22			
Tester JENC			
Recorded 08/24/22			




Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	VF3-J15,(O)T-76,BID00280,PID P361
Sample ID		
Tested 08/24/22		
Tester JENC		
Recorded 08/24/22		
SHELF LIFE OF PRODUCT	1 YEAR TIME	1 YEAR
Sample ID		
Tested 08/24/22		
Tester JENC		
Recorded 08/24/22		
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY	08/30/23
Sample ID		
Tested 08/24/22		
Tester JENC		
Recorded 08/24/22		
Actual Density	1.01593 GM/ML	1.01580 - 1.01606 1.01586
Sample ID		
Tested 08/24/22		
Tester JENC		
Recorded 08/24/22		
Lot #	2080438	
Made	08/24/22	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.


Quality Manager