



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 8638			
SODIUM HYDROXIDE 0.313 NORMAL			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 08/23/22			
Tester CNOL			
Recorded 08/23/22			
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES		YES
Sample ID			
Tested 08/23/22			
Tester CNOL			
Recorded 08/23/22			
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY		08/23/22
Sample ID			
Tested 08/23/22			
Tester CNOL			
Recorded 08/23/22			
<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0 DEG C	10.0 - 35.0	27.5
Sample ID			
Tested 08/23/22			
Tester CNOL			
Recorded 08/23/22			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY MM/DD/YYYY		08/23/22
Sample ID			
Tested 08/23/22			
Tester CNOL			
Recorded 08/23/22			
<b>AMBIENT TEST TEMPERATURE DEGC</b>	20.0 DEG C	15.0 - 30.0	20.0
Sample ID			
Tested 08/23/22			
Tester CNOL			
Recorded 08/23/22			
<b>NORMALITY TO 4 DEC PLACES</b>	0.3130 EQUIV/LITER	0.3125 - 0.3135	0.3134
Sample ID			
Tested 08/23/22			
Tester CNOL			
Recorded 08/23/22			
<b>QCSP7: STANDARDIZATION OF AQUEOUS BASES</b>	YES METHOD		YES
Sample ID			
Tested 08/23/22			
Tester CNOL			
Recorded 08/23/22			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>N.I.S.T. Traceable to SRM 84</b>	YES YES	YES
Sample ID		
Tested	08/23/22	
Tester	CNOL	
Recorded	08/23/22	
<b>APPEARANCE AND COLOR</b>	<b>CLEAR LIQUID APPEARANCE</b>	<b>PASS</b>
Sample ID		
Tested	08/23/22	
Tester	CNOL	
Recorded	08/23/22	
<b>INSTRUMENTS USED DURING PREPARATION</b>	<b>INSTRUMENT</b>	<b>M25,T72</b>
Sample ID		
Tested	08/23/22	
Tester	CNOL	
Recorded	08/23/22	
<b>EXPIRATION DATE</b>	<b>MM/DD/YY MM/DD/YYYY</b>	<b>08/30/24</b>
Sample ID		
Tested	08/23/22	
Tester	CNOL	
Recorded	08/23/22	
<b>SHELF LIFE OF PRODUCT</b>	<b>SHELF LIFE TIME</b>	<b>2 YEARS</b>
Sample ID		
Tested	08/23/22	
Tester	CNOL	
Recorded	08/23/22	
<b>Actual Density</b>	<b>1.01211 GM/ML</b>	<b>1.01199 - 1.01224 1.01217</b>
Sample ID		
Tested	08/23/22	
Tester	CNOL	
Recorded	08/23/22	
<b>Lot #</b>	<b>2080456</b>	
<b>Made</b>	<b>08/23/22</b>	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 3

Date: 08/23/22 at 1:12 PM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**