



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: M-103F			
SODIUM HYDROXIDE 0.01 NORMAL IN IPA			
Test	Target/UOM	Range	Result
Actual Normality	0.0100	0.0095 - 0.0105	0.0102
	EQ/L		
Sample ID			
Tested	08/23/22		
Tester	JENC		
Recorded	08/23/22		
PREPARED TO FORMULATION ON FILE	YES		YES
	YES		
Sample ID			
Tested	08/23/22		
Tester	JENC		
Recorded	08/23/22		
QCSP7: STANDARDIZATION OF AQUEOUS BASES	YES		YES
	METHOD		
Sample ID			
Tested	08/23/22		
Tester	JENC		
Recorded	08/23/22		
N.I.S.T. Traceable to SRM 84 (KHP)	YES		YES
	YES		
Sample ID			
Tested	08/23/22		
Tester	JENC		
Recorded	08/23/22		
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested	08/23/22		
Tester	JENC		
Recorded	08/23/22		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M55,M81,VF2-P3,M91,BID0028 0,TIDS82
Sample ID			
Tested	08/23/22		
Tester	JENC		
Recorded	08/23/22		
Actual Density	0.78595	0.78555 - 0.78635	0.78605
	GM/ML		
Sample ID			
Tested	08/23/22		
Tester	JENC		
Recorded	08/23/22		
EXPIRATION DATE	MM/DD/YYYY		02/28/23
	MM/DD/YYYY		
Sample ID			
Tested	08/23/22		
Tester	JENC		
Recorded	08/23/22		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 08/23/22 at 12:01 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Lot #	2081761
Made	08/23/22

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager