



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> FIS114F			
Chloride Standard 4.0 ppm w/v Solution			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON</b>	YES		YES
<b>FILE</b>	YES		
Sample ID			
Tested 08/26/22			
Tester MMAT			
Recorded 08/26/22			
<b>CALCULATED PPM CHLORIDE</b>	4.00	3.97 - 4.03	4.00
	PPM		
Sample ID			
Tested 08/26/22			
Tester MMAT			
Recorded 08/26/22			
<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 08/26/22			
Tester MMAT			
Recorded 08/26/22			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		P371,VF500-X27
Sample ID			
Tested 08/26/22			
Tester MMAT			
Recorded 08/26/22			
<b>EXPIRATION DATE</b>	MM/DD/YYYY		08/30/23
	MM/DD/YYYY		
Sample ID			
Tested 08/26/22			
Tester MMAT			
Recorded 08/26/22			
<b>Lot #</b>	2082632		
<b>Made</b>	08/26/22		

**Quality Manager**