



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Test	Target/UOM	Range	Result
Packaged Product: DC334F HYDROCHLORIC ACID 1.5 NORMAL IN METHANOL			
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested 08/31/22			
Tester JENC			
Recorded 08/31/22			
SAMPLE COLLECTION & HANDLING PER LAB040	YES YES		YES
Sample ID			
Tested 08/31/22			
Tester JENC			
Recorded 08/31/22			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	21.4
Sample ID			
Tested 08/31/22			
Tester JENC			
Recorded 08/31/22			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		08/31/22
Sample ID			
Tested 08/31/22			
Tester JENC			
Recorded 08/31/22			
AMBIENT TEST TEMPERATURE DEG C	20.0 DEG C	15.0 - 30.0	21.4
Sample ID			
Tested 08/31/22			
Tester JENC			
Recorded 08/31/22			
NORMALITY TO 4 DEC PLACES	1.5000 EQUIV/LITER	1.4995 - 1.5010	1.5005
Sample ID			
Tested 08/31/22			
Tester JENC			
Recorded 08/31/22			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 08/31/22			
Tester JENC			
Recorded 08/31/22			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		GC,VF12-FX,M55,P34,BID0225 3,TID411
Sample ID			
Tested 08/31/22			
Tester JENC			
Recorded 08/31/22			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 08/31/22 at 9:14 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

DATE OF SAMPLING FOR TEST

MM/DD/YYYY

08/31/22

PURPOSES

MM/DD/YYYY

Sample ID

Tested 08/31/22

Tester JENC

Recorded 08/31/22

Lot # 2081122

Made 08/31/22

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager