



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> VUL380F			
SODIUM HYDROXIDE 0.001 NORMAL IN METHANOL			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 09/08/22			
Tester JENC			
Recorded 09/08/22			
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES		YES
Sample ID			
Tested 09/08/22			
Tester JENC			
Recorded 09/08/22			
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY		09/08/22
Sample ID			
Tested 09/08/22			
Tester JENC			
Recorded 09/08/22			
<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0 DEG C	10.0 - 35.0	20.8
Sample ID			
Tested 09/08/22			
Tester JENC			
Recorded 09/08/22			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY MM/DD/YYYY		09/08/22
Sample ID			
Tested 09/08/22			
Tester JENC			
Recorded 09/08/22			
<b>AMBIENT TEST TEMPERATURE DEGC</b>	20.0 DEG C	15.0 - 30.0	20.8
Sample ID			
Tested 09/08/22			
Tester JENC			
Recorded 09/08/22			
<b>NORMALITY TO 5 DECIMAL PLACES</b>	0.00100 EQUIV/LITER	0.00090 - 0.00120	0.00098
Sample ID			
Tested 09/08/22			
Tester JENC			
Recorded 09/08/22			
<b>Actual Density</b>	0.79157 GM/ML	0.79066 - 0.79249	0.79169
Sample ID			
Tested 09/08/22			
Tester JENC			
Recorded 09/08/22			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID APPEARANCE	PASS
Sample ID		
Tested	09/08/22	
Tester	JENC	
Recorded	09/08/22	
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT	M81,VF3-J12,M55
Sample ID		
Tested	09/08/22	
Tester	JENC	
Recorded	09/08/22	
<b>EXPIRATION DATE</b>	MM/DD/YYYY	11/13/22
	MM/DD/YYYY	
Sample ID		
Tested	09/08/22	
Tester	JENC	
Recorded	09/08/22	
<b>Lot #</b>	2090217	
<b>Made</b>	09/08/22	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**