



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 6471			
NITRIC ACID 6.0 NORMAL			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 10/07/22			
Tester CCOA			
Recorded 10/07/22			
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES		YES
Sample ID			
Tested 10/07/22			
Tester CCOA			
Recorded 10/07/22			
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY		10/07/22
Sample ID			
Tested 10/07/22			
Tester CCOA			
Recorded 10/07/22			
<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0 DEG C	10.0 - 35.0	24.4
Sample ID			
Tested 10/07/22			
Tester CCOA			
Recorded 10/07/22			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY		10/07/22
Sample ID			
Tested 10/07/22			
Tester CCOA			
Recorded 10/07/22			
<b>AMBIENT TEST TEMPERATURE DEGC</b>	20.0 DEG C	15.0 - 30.0	20.9
Sample ID			
Tested 10/07/22			
Tester CCOA			
Recorded 10/07/22			
<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested 10/07/22			
Tester CCOA			
Recorded 10/07/22			
<b>QCSP5: STANDARDIZATION OF AQUEOUS ACIDS</b>	YES METHOD		YES
Sample ID			
Tested 10/07/22			
Tester CCOA			
Recorded 10/07/22			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>Actual Density</b>	1.19154 GM/ML	1.18912 - 1.19396	1.19040
Sample ID			
Tested	10/07/22		
Tester	CCOA		
Recorded	10/07/22		
<b>NORMALITY</b>	6.000 EQUIV/LITER	5.950 - 6.050	5.951
Sample ID			
Tested	10/07/22		
Tester	CCOA		
Recorded	10/07/22		
<b>N.I.S.T. Traceable to SRM 723 THAM</b>	YES YES		YES
Sample ID			
Tested	10/07/22		
Tester	CCOA		
Recorded	10/07/22		
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		GC,T-169,M55,M91
Sample ID			
Tested	10/07/22		
Tester	CCOA		
Recorded	10/07/22		
<b>EXPIRATION DATE</b>	MM/DD/YY MM/DD/YYYY		10/30/24
Sample ID			
Tested	10/07/22		
Tester	CCOA		
Recorded	10/07/22		
<b>Lot #</b>	2083149		
<b>Made</b>	10/07/22		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**