



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 1528			
BUFFER REFERENCE pH 11.0 (± 0.01)			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 10/11/22			
Tester BPHI			
Recorded 10/11/22			
Actual pH of product	11.000	10.980 - 11.020	10.992
	PH		
Sample ID			
Tested 10/11/22			
Tester BPHI			
Recorded 10/11/22			
QCSP 20 PH CALIBRATION METHODS	YES		YES
	METHOD		
Sample ID			
Tested 10/11/22			
Tester BPHI			
Recorded 10/11/22			
Actual Temp	25.0	24.0 - 26.0	24.8
	DEG C		
Sample ID			
Tested 10/11/22			
Tester BPHI			
Recorded 10/11/22			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested 10/11/22			
Tester BPHI			
Recorded 10/11/22			
INSTRUMENTS USED DURING PREPARATION	LAB EQUIPMENT		M103,GC,T-25,M106,
	INSTRUMENT		
Sample ID			
Tested 10/11/22			
Tester BPHI			
Recorded 10/11/22			
PH NIST TRACEABLE TO SRM 187	YES		YES
	YES		
Sample ID			
Tested 10/11/22			
Tester BPHI			
Recorded 10/11/22			
PH NIST TRACEABLE TO SRM 2193	YES		YES
	YES		
Sample ID			
Tested 10/11/22			
Tester BPHI			
Recorded 10/11/22			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

TRACEABLE TO NIST SRM 191I AND SRM 191II	YES YES	YES
Sample ID		
Tested	10/11/22	
Tester	BPHI	
Recorded	10/11/22	
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY	10/30/23
Sample ID		
Tested	10/11/22	
Tester	BPHI	
Recorded	10/11/22	
SHELF LIFE OF PRODUCT	SHELF LIFE TIME	1 YEAR
Sample ID		
Tested	10/11/22	
Tester	BPHI	
Recorded	10/11/22	
Lot #	2081640	
Made	10/11/22	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager