



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 9460-4L Triethanolamine 50% v/v Solution (4LG)			
Lot # 2062070			
Chemical Product: 9460 Triethanolamine 50% v/v Solution			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE		YES YES	YES
Sample ID			
Tested	06/20/22		
Tester	LBYE		
Recorded	06/20/22		
APPEARANCE		CLEAR WATER WHITE TO PALE YELLOW	PASS
Sample ID			
Tested	06/20/22		
Tester	LBYE		
Recorded	06/20/22		
INSTRUMENTS USED DURING PREPARATION		INSTRUMENT	GC,VF4-N10,
Sample ID			
Tested	06/20/22		
Tester	LBYE		
Recorded	06/20/22		
VOL/VOL%	50.0 V/V%	49.7 - 50.3	50.0
Sample ID			
Tested	06/20/22		
Tester	LBYE		
Recorded	06/20/22		
EXPIRY DATE		DD/MM/YYYY	12/30/23
Sample ID			
Tested	12/12/22		
Tester	TBUR		
Recorded	12/12/22		
Made	06/20/22		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 12/12/22 at 8:35 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager