



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 3540			
EDTA TITRANT 0.1 MOLAR (0.2N)			
Test	Target/UOM	Range	Result
Actual Molarity	0.1000 MOL/L	0.0995 - 0.1005	0.1001
Sample ID			
Tested	01/17/23		
Tester	CCOA		
Recorded	01/17/23		
QCSP4:STANDARDIZATION EDTA SOLUTIONS	YES QCSP4		YES
Sample ID			
Tested	01/17/23		
Tester	CCOA		
Recorded	01/17/23		
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested	01/17/23		
Tester	CCOA		
Recorded	01/17/23		
FORMULA WEIGHT	372.24 F.W.	= 372.24	372.24
Sample ID			
Tested	01/17/23		
Tester	CCOA		
Recorded	01/17/23		
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested	01/17/23		
Tester	CCOA		
Recorded	01/17/23		
Actual Density	1.01671 GM/ML	1.01605 - 1.01736	1.01666
Sample ID			
Tested	01/17/23		
Tester	CCOA		
Recorded	01/17/23		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		,101,T-9,M55,BID00488,PIDM9 1
Sample ID			
Tested	01/17/23		
Tester	CCOA		
Recorded	01/17/23		
Traceable to N.I.S.T. SRM 915	YES YES		YES
Sample ID			
Tested	01/17/23		
Tester	CCOA		
Recorded	01/17/23		



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

EXPIRATION DATE	MM/DD/YY MM/DD/YYYY	01/30/25
Sample ID		
Tested	01/17/23	
Tester	CCOA	
Recorded	01/17/23	
SHELF LIFE OF PRODUCT	2 YEARS TIME	2 YEARS
Sample ID		
Tested	01/17/23	
Tester	CCOA	
Recorded	01/17/23	
Lot #	2123069	
Made	01/17/23	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager