



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: FIS109F			
Nitrate Standard 2.0 ppm as NO3			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 02/28/23			
Tester MMAT			
Recorded 02/28/23			
CALCULATED PPM NITRATE (NO3)	2.00	1.95 - 2.05	2.00
	PPM W/V		
Sample ID			
Tested 02/28/23			
Tester MMAT			
Recorded 02/28/23			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 02/28/23			
Tester MMAT			
Recorded 02/28/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P394,VF1-Z21
Sample ID			
Tested 02/28/23			
Tester MMAT			
Recorded 02/28/23			
EXPIRATION DATE	MM/DD/YYYY		05/03/23
	MM/DD/YYYY		
Sample ID			
Tested 02/28/23			
Tester MMAT			
Recorded 02/28/23			
Lot #	3022777		
Made	02/28/23		

Richard [Signature]
Quality Manager