



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

Test	Target/UOM	Range	Result
<b>Packaged Product:</b> CPA004F			
	SODIUM THIOSULFATE 0.1576 NORMAL		
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 03/06/23			
Tester DTRA			
Recorded 03/06/23			
<b>Actual Normality</b>	0.1576 EQ/L	0.1571 - 0.1581	0.1580
Sample ID			
Tested 03/06/23			
Tester DTRA			
Recorded 03/06/23			
<b>Actual Density</b>	1.01861 GM/ML	1.01820 - 1.01902	1.01860
Sample ID			
Tested 03/06/23			
Tester DTRA			
Recorded 03/06/23			
<b>QCSP10:STANDARDIZATION OF SODIUM THIOSUL</b>	YES QCSP10		YES
Sample ID			
Tested 03/06/23			
Tester DTRA			
Recorded 03/06/23			
<b>QCSP35:STANDARDIZATION OF SODIUM THIOSUL</b>	YES QCSP35		YES
Sample ID			
Tested 03/06/23			
Tester DTRA			
Recorded 03/06/23			
<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 03/06/23			
Tester DTRA			
Recorded 03/06/23			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		T-28,M112,BID-00787,PID-M81
Sample ID			
Tested 03/06/23			
Tester DTRA			
Recorded 03/06/23			
<b>SHELF LIFE OF PRODUCT</b>	6 MONTHS TIME		6 MONTHS
Sample ID			
Tested 03/06/23			
Tester DTRA			
Recorded 03/06/23			



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 2

Date: 03/06/23 at 10:12 AM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

**EXPIRATION DATE**

MM/DD/YYYY

09/30/23

MM/DD/YYYY

Sample ID

Tested 03/06/23

Tester DTRA

Recorded 03/06/23

**Lot #** 3010569

**Made** 03/06/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**