



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: FIS111F			
Sulfate Standard 10.0 ppm w/v Solution			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 03/10/23			
Tester MMAT			
Recorded 03/10/23			
CALCULATED PPM SULFATE (SO4)	10.00	9.95 - 10.05	10.00
	PPM		
Sample ID			
Tested 03/10/23			
Tester MMAT			
Recorded 03/10/23			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 03/10/23			
Tester MMAT			
Recorded 03/10/23			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P83,VF500-X30
Sample ID			
Tested 03/10/23			
Tester MMAT			
Recorded 03/10/23			
EXPIRATION DATE	MM/DD/YYYY		09/30/23
	MM/DD/YYYY		
Sample ID			
Tested 03/10/23			
Tester MMAT			
Recorded 03/10/23			
Lot #	3031061		
Made	03/10/23		

Richard [Signature]

Quality Manager