



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Test	Target/UOM	Range	Result
Packaged Product: ISP162F			
	SULFURIC ACID MOBILE PHASE 0.001 NORMAL		
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 03/27/23			
Tester BPHI			
Recorded 03/27/23			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 03/27/23			
Tester BPHI			
Recorded 03/27/23			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		03/27/23
Sample ID			
Tested 03/27/23			
Tester BPHI			
Recorded 03/27/23			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	21.8
Sample ID			
Tested 03/27/23			
Tester BPHI			
Recorded 03/27/23			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		03/27/23
Sample ID			
Tested 03/27/23			
Tester BPHI			
Recorded 03/27/23			
AMBIENT TEST TEMPERATURE DEGC	20.0 DEG C	15.0 - 30.0	21.8
Sample ID			
Tested 03/27/23			
Tester BPHI			
Recorded 03/27/23			
NORMALITY TO 5 DECIMAL PLACES	0.00100 EQUIV/LITER	0.00090 - 0.00115	0.00099
Sample ID			
Tested 03/27/23			
Tester BPHI			
Recorded 03/27/23			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 03/27/23			
Tester CNOL			
Recorded 03/27/23			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Final Product Filtered	YES	YES
	YES	
Sample ID		
Tested	03/27/23	
Tester	CNOL	
Recorded	03/27/23	
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	GC,T-225,HEDPAK,M55,P109
Sample ID		
Tested	03/27/23	
Tester	CNOL	
Recorded	03/27/23	
SHELF LIFE OF PRODUCT	6 MONTHS	6 MONTHS
	TIME	
Sample ID		
Tested	03/27/23	
Tester	BPHI	
Recorded	03/27/23	
EXPIRATION DATE	MM/DD/YYYY	09/30/23
	MM/DD/YYYY	
Sample ID		
Tested	03/27/23	
Tester	BPHI	
Recorded	03/27/23	
Actual Density	0.99825	0.99811 - 0.99839
	GM/ML	0.99824
Sample ID		
Tested	03/27/23	
Tester	BPHI	
Recorded	03/27/23	
Lot #	3032330	
Made	03/27/23	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Richard [Signature]

Quality Manager