



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 8670			
SODIUM HYDROXIDE 10.0 NORMAL			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 03/28/23			
Tester JENC			
Recorded 03/28/23			
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES		YES
Sample ID			
Tested 03/28/23			
Tester JENC			
Recorded 03/28/23			
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY		03/28/23
Sample ID			
Tested 03/28/23			
Tester JENC			
Recorded 03/28/23			
<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0 DEG C	10.0 - 35.0	21.4
Sample ID			
Tested 03/28/23			
Tester JENC			
Recorded 03/28/23			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY MM/DD/YYYY		03/28/23
Sample ID			
Tested 03/28/23			
Tester JENC			
Recorded 03/28/23			
<b>AMBIENT TEST TEMPERATURE DEGC</b>	20.0 DEG C	15.0 - 30.0	21.4
Sample ID			
Tested 03/28/23			
Tester JENC			
Recorded 03/28/23			
<b>NORMALITY</b>	10.000 EQUIV/LITER	9.950 - 10.050	10.039
Sample ID			
Tested 03/28/23			
Tester JENC			
Recorded 03/28/23			
<b>QCSP7: STANDARDIZATION OF AQUEOUS BASES</b>	YES METHOD		YES
Sample ID			
Tested 03/28/23			
Tester JENC			
Recorded 03/28/23			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>N.I.S.T. Traceable to SRM 84 (KHP)</b>	YES		YES
	YES		
Sample ID			
Tested	03/28/23		
Tester	JENC		
Recorded	03/28/23		
<b>APPEARANCE AND COLOR</b>	<b>CLEAR WATER WHITE</b>		<b>PASS</b>
	<b>LIQUID</b>		
	<b>APPEARANCE</b>		
Sample ID			
Tested	03/28/23		
Tester	JENC		
Recorded	03/28/23		
<b>INSTRUMENTS USED DURING PREPARATION</b>	<b>INSTRUMENT</b>		<b>M25,T-72,BID-02253</b>
Sample ID			
Tested	03/28/23		
Tester	JENC		
Recorded	03/28/23		
<b>EXPIRATION DATE</b>	<b>MM/DD/YY</b>		<b>03/30/25</b>
	<b>MM/DD/YYYY</b>		
Sample ID			
Tested	03/28/23		
Tester	JENC		
Recorded	03/28/23		
<b>SHELF LIFE OF PRODUCT</b>	<b>SHELF LIFE</b>		<b>2YEARS</b>
	<b>TIME</b>		
Sample ID			
Tested	03/28/23		
Tester	JENC		
Recorded	03/28/23		
<b>Actual Density</b>	<b>1.33032</b>	<b>1.32653 - 1.33412</b>	<b>1.33017</b>
	<b>GM/ML</b>		
Sample ID			
Tested	03/28/23		
Tester	JENC		
Recorded	03/28/23		
<b>Lot #</b>	<b>3020973</b>		
<b>Made</b>	<b>03/28/23</b>		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 3

Date: 03/28/23 at 10:23 AM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**