04/26/23

Made

Page: 1

Date: 04/26/23 at 1:58 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

EP043F **Packaged Product:** CALCIUM CHLORIDE 2.0 MOLAR pH 1.0 W/HCl Target/UOM Result Range Test PREPARED TO FORMULATION ON YES YES YES **FILE** Sample ID Tested 04/26/23 Tester **LWIL** Recorded 04/26/23 Actual pH of product 0.500 - 1.010 0.995 0.750 PH Sample ID Tested 04/26/23 Tester **LWIL** Recorded 04/26/23 2.00 **CALCULATED MOLARITY** 1.98 - 2.02 2.00 MOL/L Sample ID Tested 04/26/23 Tester **LWIL** Recorded 04/26/23 INSTRUMENTS USED DURING **INSTRUMENT** M113,T-23 **PREPARATION** Sample ID Tested 04/26/23 Tester **LWIL** Recorded 04/26/23 3042578 Lot#

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager