

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: BAS711F HYDROCHLORIC ACID 0.001 NORMAL (STANDARDIZED)			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 04/24/23			
Tester JENC			
Recorded 04/24/23			
SAMPLE COLLECTION & HANDLING	YES		YES
PER LAB040	YES		
Sample ID			
Tested 04/24/23			
Tester JENC			
Recorded 04/24/23			
DATE OF SAMPLING FOR TEST	MM/DD/YY		04/24/23
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested 04/24/23			
Tester JENC			
Recorded 04/24/23			
AMBIENT TEMPERATURE DURING	25.0	10.0 - 35.0	20.8
SAMPLING DEGC	DEG C		
Sample ID			
Tested 04/24/23			
Tester JENC			
Recorded 04/24/23			
DATE OF TESTING OF SAMPLE	MM/DD/YY		04/24/23
	MM/DD/YYYY		
Sample ID			
Tested 04/24/23			
Tester JENC			
Recorded 04/24/23			
AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	20.8
DEG C	DEG C		
Sample ID			
Tested 04/24/23			
Tester JENC			
Recorded 04/24/23			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested 04/24/23			
Tester JENC			
Recorded 04/24/23			
INSTRUMENTS USED DURING	INSTRUMENT		VF200-T6,T-158,BID-02253,P1
PREPARATION			09
Sample ID			
Tested 04/24/23			
Tester JENC			
Recorded 04/24/23			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 04/24/23 at 4:16 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

NORMALITY TO 5 DECIMAL PLACES

0.00100
EQUIV/LITER

0.00080 - 0.00120

0.00103

Sample ID
Tested 04/24/23
Tester JENC
Recorded 04/24/23

Actual Density

0.99825
GM/ML

0.99804 - 0.99846

0.99823

Sample ID
Tested 04/24/23
Tester JENC
Recorded 04/24/23

EXPIRATION DATE

MM/DD/YY
MM/DD/YYYY

10/30/23

Sample ID
Tested 04/24/23
Tester JENC
Recorded 04/24/23

Lot # 3040571

Made 04/24/23

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager